Certificate of Analysis

Conductivity Calibration Solution
3169

1725 Brannum Lane
Yellow Springs, Ohio
45387-1107 USA
+1 937-767-7241

Calibration Equipment:

- IET 1693 RLC Digibridge
  Instrument #: 7120
  Calibration Date: 05/31/2026
- Agilent 34401A 6½ Digit Multimeter
  Instrument #: 6859
  Calibration Date: 11/30/2026

Testing Results

Lot Number: 26D100672

Nominal Conductivity (microsiemens/cm):	50K
Specification Range (microsiemens/cm):	49500 - 50500
Test Result (microsiemens/cm):	49668

Traceability:
We certify that we have verified this product using reference solutions, equipment, and standards that are traceable to the National Institute of Standards and Technology (NIST) or are derived from accepted values of natural physical constants. As manufactured, the product meets or exceeds its current specifications.

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.

Technician	Date
Joey Flores	4/29/2026