# *Certificate of Analysis*

## Conductivity Calibration Solution
## 3167

1725 Brannum Lane
Yellow Springs, Ohio
45387-1107 USA
+1 937-767-7241

Calibration Equipment:

- IET 1693 RLC Digibridge
  Instrument #: 7120
  Calibration Date: 05/31/2026
- Agilent 34401A 6½ Digit Multimeter
  Instrument #: 6859
  Calibration Date: 11/30/2026

## Testing Results

Lot Number: 26D100674

| | |
|---|---|
| Nominal Conductivity (microsiemens/cm): | 1K |
| Specification Range (microsiemens/cm): | 990 - 1010 |
| Test Result (microsiemens/cm): | 995 |

Traceability:
We certify that we have verified this product using reference solutions, equipment, and standards that are traceable to the National Institute of Standards and Technology (NIST) or are derived from accepted values of natural physical constants. As manufactured, the product meets or exceeds its current specifications.

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.

| Technician | Date |
|---|---|
| Joey Flores | 5/4/2026 |